

A Review paper of the spectroscopy

Author:- Puja Vijay Kardile
Guide:- Miss. Prof. Komal Baride
Principal:- DR. Sanjay Ingle

Dharmraj Shaikshnaik Pratishthan College Of Pharmacy, Ahilyanagr, 414006, Maharashtra, India.

Date of Submission: 01-10-2024

Date of Acceptance: 10-10-2024

ABSTRACT:

One scientific method for examining how matter and electromagnetic waves interact is spectroscopy. By examining how different substances absorb, emit, or scatter light, one can gain insight into their composition, structure, and qualities.

The study of how matter absorbs and emits light and other radiations is known as spectroscopy. Similar to how a prism splits light into a rainbow, it entails dividing light, or electromagnetic radiation, into its component wavelengths, or spectrum.

One significant and cutting-edge analytical tool employed in the pharmaceutical business over the past 35 years is ultraviolet spectroscopy. The analytical technique measures the amount of monochromatic light absorbed by colourless substances in the near ultraviolet (200–400) range. The processes required to ascertain the "identity, strength, quality and purity" of such chemicals are included in the pharmaceutical analysis. It also covers the examination of raw materials and intermediates used in the pharmaceutical production process. A spectrophotometer covering the ultraviolet range operates on the basic principle of light passing through a solvent-filled cell and onto a photoelectric cell, which converts radiant energy into electrical energy that is measured by a galvanometer.


Keywords : Spectroscopy , Ultra Violent , Identity , Strength , Quality , Purity , Galvanometer.

Application :

Spectroscopy is used in many fields, including:
Chemistry: Identifying chemical compounds and studying reaction mechanisms.
Physics: Investigating atomic and molecular structures.
Astronomy: Analyzing the composition of stars and galaxies.
Medicine : Diagnosing diseases and monitoring biological processes.

I. INTRODUCTION :

The study of the electromagnetic spectrum that is created when matter interacts with one another or emits electromagnetic radiation is known as spectroscopy. Every chemical element and compound has an own characteristic spectrum and may absorb and disperse light over a specific range of frequencies or wavelengths. To put it simply, spectroscopy is a technique for determining the intensity of light that passes through a substance and the amount of light that is absorbed by it.



Six fundamental Types of spectroscopy exist.

- 1) IR spectroscopy
- 2) Ultra Violent spectroscopy
- 3) Nuclear magnetic resonance (NMR)
- 4) Mass spectroscopy
- 5) Raman spectroscopy
- 6) Absorption Spectroscopy

Infrared spectroscopy, also referred to as vibrational spectroscopy or infrared spectroscopy, is the measurement of the absorption, emission, or reflection of infrared light by matter. Its use entails analyzing and identifying functional groups or chemical compounds in gaseous, liquid, or solid phases.

1) Infrared Spectroscopy (IR):

IR is the most potent analytical method that allows for chemical identification.

This method is based only on the observation that a chemical material exhibits pronounced selective absorption in the infrared spectrum.

Infrared spectroscopy, also referred to as vibrational spectroscopy or infrared spectroscopy, is the measurement of the absorption, emission, or reflection of infrared light by matter. Its use entails analyzing and identifying functional groups or chemical compounds in gaseous, liquid, or solid phases. The process or technique of infrared spectroscopy is carried out using an infrared spectrometer. An infrared spectrum is produced by it. An IR spectrum can be shown on a graph by placing infrared light transmittance (or absorbance) on the vertical axis and frequency or wavelength on the horizontal axis.

Reciprocal centimetre, often known as wave numbers, are commonly used units of frequency in infrared spectra. They are denoted by the notation cm^{-1} . Infrared wavelength units are often represented by the symbol μm , or micrometre (previously known as "microns).


Principle: The vibrational and rotational energy of a molecule is related to infrared spectroscopy. The molecule absorbs infrared light when the frequency of the radiation is the same as the natural frequency of vibration.

Molecules are excited from a lower vibrational state to a higher one when infrared radiation is absorbed. Numerous closely spaced rotational levels are connected to each vibrational level. As a result, "vibrational-rotational spectroscopy" is another name for IR spectroscopy.



Construction: Infrared (IR) radiation is the general term for the region of the electromagnetic spectrum that is between the visible and microwave domains. It is the small range of $4000\text{--}400\text{ cm}^{-1}$ that is most

helpful to practicing organic chemists. There has been a lot of interest in the near-IR, in addition to the far-IR range ($700\text{--}200\text{ cm}^{-1}$). Even if the molecule as a whole is represented in the IR spectrum, specific atomic groups give rise to bands at or near the same frequency with little consideration to the remainder of the molecule's structure. The enduring nature of these traits indicates that allows the chemist to consult generalized charts of characteristics group frequencies and perform a quick inspection to gain valuable structural information.



Instrumentation: Radiation Source: The infrared radiation coming from the source needs to be constant and strong. Typical sources consist of:

- o Nernst glower
- o An incandescent lightbulb
- o The Mercury Arc
- o Tungsten light
- o Global source
- o Nickel wire

2. Sample Cells and Sampling: The sample for analysis is kept in the sample cells. Depending on whether the sample is in a liquid, gas, or solid state, different techniques are applied.

3. Monochromators: These apparatuses pick out particular infrared radiation frequencies. They can be prisms or diffraction gratings that disperse the IR light into its component wavelengths.

4. Detectors: Following the radiation's passage through the sample, detectors gauge its intensity. Typical kinds consist of:

- o Thermocouples
- o Bolometers
- o Detectors pyroelectric
- o Detectors that are photoconductive

5. Fourier Transform Infrared (FTIR) Spectrometers: Fourier transform methods are frequently employed by contemporary infrared spectrometers. o An interferometer (equipped with

a beam splitter, moving mirror, and fixed mirror) is a component of an FTIR spectrometer.

- o A detector
 - o A computer to create the spectrum and carry out the Fourier transform
6. Records: The IR spectrum, which is a plot of IR intensity against wavelength or frequency, is recorded by these sensors.

Benefits :

Infrared (IR) spectroscopy is a useful instrument in a variety of fields since it has multiple benefits.

1. Non-Destructive: Since IR spectroscopy is a non-destructive method, the sample is not damaged and can be utilized for additional examination.
2. Minimal Sample Preparation: Preparing samples is usually minimal or nonexistent, which streamlines the procedure and lowers the chance of contamination.
3. High Speed: The method enables quick data collecting and frequently yields results in a matter of seconds.
4. Versatility: It is capable of analyzing a variety of sample forms, such as gases, liquids, and solids.
5. High Resolution: Accurate compound identification and characterization are made possible by the high resolution spectra that IR spectroscopy offers.
6. Quantitative and Qualitative Analysis: It can be applied to quantitatively quantify the concentrations of chemicals as well as qualitatively identify them.

Drawbacks:

- 1) Limited Structural Information: The main use of IR spectroscopy is to identify functional groups; it is not a reliable source of information on the general structure of molecules.
- 2) Complex Spectra: It can be difficult to precisely identify every component in samples with several components or overlapping spectra.
- 3) Sensitivity to Water: The analysis may be hampered by water's significant IR radiation absorption. Before analysis, the samples must be completely dried.
- 4) Sample Preparation: Sample preparation can take a lot of time and expertise, especially when making thin films.
- 5) spectrum Interference: Inaccurate results may result from spectrum interference caused by other molecules or contaminants in the sample.
- 6) Quantitative Restrictions: Although infrared spectroscopy is a useful tool for quantitative analysis, its applicability for accurate quantitative measurements may be limited by several issues.

Application:

Structural clarification and functional group identification.

Identifying a pharmacological substance's impurities.

polymer research.

Measurement and Evaluation

Finding the Cis-Trans isomer.

Bottom of the Form.

2) Ultra Violent Spectroscopy :

The ultra violet spectroscopy, also known as ultra-violet spectroscopy, An analytical method called UV-Visible (UV-Vis) spectroscopy is used to determine how much visible and ultraviolet (UV) light a sample absorbs. By using their characteristics related to light absorption, these substances can be identified and quantified.



Principle:


The basis of UV-Vis spectroscopy is light absorption in the visible and UV portions of the electromagnetic spectrum. A sample's molecules absorb certain wavelengths of light, which results in electronic transitions from the ground state to the excited state.

In chemistry and biology, ultraviolet-visible (UV-Vis) spectroscopy is a commonly used analytical method for identifying and measuring molecules. Here's a thorough rundown:

Parts of a UV-Vis Spectrophotometer

- 1) Light Source: Usually a tungsten lamp for visible light and a deuterium lamp for ultraviolet light.
- 2) Monochromator: Chooses particular light wavelengths to let through the material.
- 3) Sample Holder: A cuvette is typically used to hold the sample solution.

4) Detector: Determines the transmitted light's intensity and produces an electrical signal from it.



Laws :

The Beer-Lambert Law

The Beer-Lambert Law, which relates absorbance (A) to the concentration (C) of the absorbing species, route length (b), and molar absorptivity (ϵ), is the fundamental idea underlying UV-Vis spectroscopy.

$$A = \epsilon bc$$

Type of transition: $\pi \rightarrow \pi^*$: Transitions between orbitals that are pi anti-bonding and pi bonding. Orbitals $n \rightarrow \pi^*$ change from non-bonding to pi anti-bonding.

$\sigma \rightarrow \sigma^*$: Changes in orbital states from sigma anti-bonding to sigma bonding.

$n \rightarrow \sigma^*$: Orbitals that go from non-bonding to sigma anti-bonding.

Benefits :

1) Non-Destructive: After analysis, the sample is left intact, enabling additional testing if necessary.

2) Speedy and Easy: The method is simple to apply and allows for speedy measurement taking.

3) High Sensitivity: Able to identify analyte concentrations at low levels.

Analyses, both quantitative and qualitative, are helpful in identifying chemicals and figuring out their quantities.

4) Broad Range of Uses: Relevant to industries such as biochemistry, environmental science, pharmaceuticals, and the food and beverage sector.

5) Minimal Sample Preparation: This technique often saves time and resources by requiring little to no sample preparation.

6) Cost-effective: Usually less expensive than other analytical methods like mass spectrometry or NMR.

7) Versatility: Able to evaluate liquids, solids, and gases, among other sample types.

Real-time monitoring is helpful for tracking changes in concentration over time and reaction kinetic.

Drawbacks :

1) Restricted to Chromophores: Analysis is only possible for compounds containing chromophores, or molecules that absorb UV or visible light.

2) Interference: Inaccurate results may arise from other absorbing species in the sample interfering with the test.

3) Low Specificity: It might not be able to differentiate apart substances with comparable absorption spectra.

4) Sample Clarity: Light can be scattered by turbid or brightly colored samples, which might compromise measurement accuracy.

5) Concentration Range: Samples must be diluted since extremely high concentrations may cause the Beer-Lambert Law to be violated.

6) Baseline Drift: The accuracy of absorbance values can be impacted by baseline drift, which is caused by instrumental factors.

7) Effects of Solvent: The absorption spectra may be affected by the solvent selected, which could make the analysis more difficult.

8) Temperature Sensitivity: Variations in temperature can have an impact on absorbance, therefore measurements must be carefully controlled.

9) Restricted Structural Information: UV-Vis spectroscopy, in contrast to methods like NMR or mass spectrometry, offers only a limited amount of structural information.

Application :

1) Quantitative analysis: figuring out how much of a material is in a solution.

2) Qualitative Analysis: Compound identification using absorption spectra.

3) Reaction Monitoring: Tracking variations in absorbance to examine the kinetics of reactions.

4) Analyzing proteins, nucleic acids, and other macromolecules is known as biological studies.

Strength and limitations

Strength:

1) Non-harmful

2) Quick & easy

3) Appropriate for a variety of chemicals.

Strength and limitations :

Limitations:

- 1) Needs chromophores, which are molecules that absorb UV-visible light.
- 2) Absorbing species in the sample causing interference.

Typical Uses :

- 1) Pharmaceuticals: Calculating the concentrations of drugs.

Environmental testing involves measuring air and water pollutants. Food and Beverage Industry: Examining additive concentration and colour .

- 3) Nuclear Magnetic Resonance :

NMR spectroscopy: Spectroscopy using Nuclear Magnetic Resonance (NMR)

It is an effective analytical method for figuring out a molecule's dynamics, structure, reaction state, and chemical environment.

NMR Spectroscopy

Principle: The foundation of NMR spectroscopy is the radiofrequency energy that nuclei in a magnetic field absorb. Non-zero spin nuclei, such as carbon-13 (^{13}C) and hydrogen-1 (^1H), orient themselves either in front of or away from an external magnetic field. These nuclei absorb energy and change their spin states in response to radiofrequency pulses. In order to learn more about the molecular structure, the energy that is released as they return to their initial condition is detected and examined.

Components of an NMR spectrophotometer:

- 1) Magnet: Usually superconducting magnets, these provide a powerful and consistent magnetic field.
- 2) Radiofrequency Transmitter: Produces the pulses of radiofrequency energy necessary to excite nuclei.
- 3) Sample Holder: The sample is often placed in a glass tube.
- 4) Detector: Calculates the radiofrequency signals that the sample emits.
- 5) Computer: Creates the NMR spectrum by processing the data.

Types of NMR spectroscopy:

Hydrogen atoms within a molecule are examined using proton NMR (^1H NMR).

The carbon atoms of a molecule are examined using carbon-13 NMR, or ^{13}C NMR.

Other Nuclei: Other nuclei such as nitrogen-15 (^{15}N), phosphorus-31 (^{31}P), and fluorine-19 (^{19}F) can also be studied using NMR.

Strength and limitations:

Strength: 1) Non-Destructive: Following analysis, the material is left intact.

2) Detailed Structural Information: Offers in-depth details regarding the dynamics and structure of molecules.

3) Versatile: Suitable for a variety of samples, such as gases, liquids, and solids.

Limitations:

1) Sensitivity: Compared to other methods like mass spectrometry, this method requires comparatively large amounts of sample.

2) Cost: The equipment and maintenance have a high initial cost.

3) Complexity: Needs knowledgeable operators and intricate data processing.

Common use:

1) Pharmaceuticals: Measuring and classifying medicinal ingredients.

Determining the structure of novel substances through research in chemistry.

3) Biochemistry: the study of biomolecules' composition and actions.

Application:

Structural Clarification: Establishing an Organic Compound's Structure

Quantitative Analysis: Determining the chemical concentration.

Investigating the motion and interactions of molecules is the study of molecular dynamics.


Biological Studies: Examining biomolecules such as proteins and nucleic acids.

4) Raman Spectroscopy :

An effective analytical method for comprehending a system's rotational, vibrational, and other low-frequency modes is Raman spectroscopy. Here are some essential specifics.

Principle:

Raman scattering, or the inelastic scattering of photons, is the basis for Raman spectroscopy. The majority of photons in a monochromatic light beam—typically produced by a laser—are elastically scattered when they interact with molecules. But a tiny percentage of the light is scattered at various energies, revealing details about the molecular vibrations.



Process:

- 1) Illumination: A laser beam is used to illuminate a sample.
- 2) Scattering: The energy of the laser photons changes as a result of the light's interaction with phonons, molecules, or other excitations.

Detection: The dispersed light is gathered and examined to identify the energy shifts that correlate to particular vibrations of the molecules.

Laws:

The fundamental concepts and selection rules that dictate the mode and timing of Raman scattering are the foundation of Raman spectroscopy. Here are a few important rules and guidelines:

Modes:

1) Raman Effect

The Raman effect is the inelastic scattering of photons by molecules, resulting in a shift in the energy of the scattered photons. This shift provides information about the vibrational modes of the molecules.

2. Selection Rules

For a molecule to exhibit a Raman effect, there must be a change in its polarizability during vibration. This means that the vibrational mode must cause a change in the distribution of the electron cloud around the molecule.

3. Stokes and Anti-Stokes Scattering

Stokes Scattering: Occurs when the scattered photon has less energy than the incident photon. This is more common because most molecules are in their ground vibrational state at room temperature.

1. The Raman Effect

The inelastic scattering of photons by molecules, which causes a change in the scattered photons' energy, is known as the Raman effect. This shift reveals details about the molecular vibrational modes.

2. Guidelines for Selection

A molecule's polarizability must alter during vibration in order for it to display a Raman effect. This indicates that the distribution of the electron cloud surrounding the molecule must alter as a result of the vibrational mode.

3. Dispersal of Stokes and Anti-Stokes

When the scattered photon's energy is lower than the incident photon's, Stokes Scattering takes place. Because most molecules at ambient temperature are in their ground vibrational state, this is more common.

4. Rayleigh Scattering: Rayleigh scattering, in which the scattered photons have the same energy as the incident photons, accounts for the majority of dispersed light. In order to concentrate on the inelastic scattering, this is filtered out in Raman spectroscopy.

5. Polarizability: The relationship between the intensity of Raman scattering and the polarizability of the molecule is proportional. Greater polarizability variations produce more intense Raman emissions.

6. Raman Resonance Effect:

Raman scattering is significantly increased when the incident light frequency approaches the molecule's electronic transition. This is referred to as the resonance Raman effect, and it can make the method more sensitive.

Benefits:

Non-Destructive: Doesn't cause any harm to the specimen. A minimal amount of sample preparation is necessary, if any at all.

High Specificity: Able to deliver precise molecular data.

limitations:

Fluorescence Interference: Raman signals may be affected by fluorescent materials.

Weak Signal: Due to the intrinsic weakness of Raman scattering, sensitive detecting tools are needed.

Drawbacks :

1. Fluorescence Interference: It might be challenging to get good spectra when fluorescent samples obscure the Raman signal.
2. Weak Signal: Due to the intrinsic weakness of Raman scattering, expensive and extremely sensitive detection equipment is frequently needed.
3. Sample Heating: A high laser power may heat the sample locally, which may change or harm it.
4. Complicated Data Interpretation: Accurate interpretation of the spectra can be challenging, requiring specialized knowledge and sophisticated tools.
5. Limited Sensitivity for Low Concentrations: The weak Raman signal makes it difficult to detect low concentrations of a chemical.

Application:

1. Chemical analysis: Involves determining the structure and composition of molecules.
2. Material science: is the study of material characteristics, such as phase and Crystallinity .
3. Pharmaceuticals: Analysis of drug formulation and quality control.
4. Identifying compounds in forensic investigations is known as forensics.
- 5) Mass spectroscopy:
The analytical method of mass spectrometry (MS) is used to calculate the mass-to-charge ratio of ions. The following are some specifics of mass spectrometry:

Principle: The basic idea behind mass spectrometry is to measure the mass-to-charge ratios of charged molecules or fragments of molecules produced by ionizing chemical substances. The findings are displayed as a mass spectrum, which is a plot of the mass-to-charge ratio against the ion signal.



Components:

Ion Source: Produces ions from the sample. Electrospray Ionization (ESI) and Electron Ionization (EI) are common techniques.

The mass-to-charge ratio of the ions .
mass analyser : to separate them. There are three types: Ion Trap, Time-of-Flight (TOF), and Quadrupole.

detector: finds the ions and sends information to create the mass spectrum. Often utilized are electron multipliers

Process:

Sample Introduction: There are three ways to introduce a sample: as a solid, liquid, or gas.

Ionization: To create charged particles, the sample is ionized.

Mass analysis: The mass-to-charge ratio is used to separate the ions.

Laws:

A number of fundamental ideas and laws governing how ions behave in electric and magnetic fields form the basis of mass spectrometry's operations. Here are a few important rules and guidelines:

1. De-ionization/ionization

Ionization is the process of changing neutral molecules into charged ions. Numerous techniques, including Electrospray Ionization (ESI), Electron Ionization (EI), and Matrix-Assisted Laser Desorption/Ionization (MALDI), can be used to accomplish this.

2. The ratio of mass to charge (m/z)

The mass-to-charge ratio (m/z) of ions is determined by the mass spectrometer. Because it dictates how ions will be separated and identified in the mass analyzer, this ratio is very important.

3. The Lorentz Force Law

The Lorentz force law, which states that the force acting on a charged particle in an electric and magnetic field is given by: $F = q(E + v \times B)$ where (\mathbf{F}) is the force, (q) is the particle's charge, (\mathbf{E}) is the electric field, (\mathbf{v}) is the particle's velocity, and (\mathbf{B}) is the magnetic field, governs the motion of ions in a mass spectrometer.

4. The Principle of Time-of-Flight (TOF)

Ions are accelerated to the same kinetic energy by an electric field in Time-of-Flight mass spectrometry. The ions' mass-to-charge ratio squared is used to calculate the amount of time it takes for them to get at the detector.

The formula is $t = m/z^2Vd$, where V is the accelerating voltage, d is the distance travelled m/z is the mass-to-charge ratio, and t is the time of flight.

5. Mass Filter Quadrupole

Based on the mass-to-charge ratio of the ions, a quadrupole mass filter selectively stabilizes or destabilizes their trajectories using oscillating electric fields. At any given time, only ions with a particular m/z ratio will be able to flow past the filter.

6. Isotopic Proliferation

The isotopic composition of an element in a sample can be ascertained using mass spectrometry. The average atomic mass and the presence of certain isotopes are determined using the relative abundance of isotopes.

Detection: A mass spectrum is produced using the data obtained from the detection of the separated ions.

Application:

1) Chemical analysis: identifying unidentified substances and ascertaining the make-up and structure of molecules.

Analyzing proteins and peptides is known as proteomics.

3) Pharmaceuticals: Quality assurance and medication development.

4) Environmental Analysis: Pollution and Contamination Detection

Benefits :

°High Sensitivity: Able to identify compounds at extremely low quantities.

°Specificity: Offers comprehensive molecular details.

°Variability: Fits a variety of samples and substances.

Drawbacks :

Complexity: To operate and interpret the data accurately, the procedure need for specialized equipment and a high degree of knowledge.

Cost: Some laboratories may find it prohibitive to purchase and maintain mass spectrometers because of the high initial cost.

Sample Preparation: Preparing some samples in-depth can take a lot of time and increase the risk of contamination.

Matrix Effects: The accuracy and ionization efficiency of the results can be impacted by the presence of other chemicals in the sample.

Restricted Quantitative Analysis: Although mass spectrometry is a great tool for qualitative analysis, differences in ionization efficiency can make quantitative analysis difficult.

Limitations:

Complexity: Needs specialized knowledge and equipment.

Cost: High upkeep and upfront investment expenses.

Sample Preparation: A thorough preparation may be necessary for some samples.

6) Absorption Spectroscopy

Absorption spectroscopy: The method of measuring an object's absorption of electromagnetic radiation is called absorption spectroscopy.

The following are some specifics of absorption spectroscopy:

Principle:

The foundation of absorption spectroscopy is the idea that certain wavelengths of light are absorbed by molecules. A sample absorbs some wavelengths of light and transmits the remaining light when it passes through it. To generate an absorption spectrum, the amount of light absorbed at each wavelength is measured.

Components:

1. Light Source: This source emits electromagnetic radiation, which may be visible, infrared, or ultraviolet in nature.

2. Sample Holder: Holds the sample that allows light to flow through.

3. Monochromator: Isolates specific wavelengths of light.

4. Detector: Measures the intensity of transmitted light at different wavelengths.

Process:

1. Illumination: The source's light falls on the sample.

The sample's ability to absorb light at particular wavelengths that correlate to the energy differences between molecular or atomic states is known as absorption.

2. Detection: An absorption spectra is collected once the transmitted light has been identified.

Types of Absorption Spectroscopy:

UV-Vis Spectroscopy: Analyzes samples using both visible and ultraviolet light. widely employed in the investigation of electronic transitions in molecules.

Infrared (IR) Spectroscopy: Investigates molecular vibrational transitions using infrared light.

X-ray Absorption Spectroscopy: This technique makes use of X-rays to examine molecules' and atoms' electronic structures.

Laws:

The Beer-Lambert Law

Beer's Law, sometimes referred to as Beer-Lambert Law, establishes a relationship between light absorption and the characteristics of the medium through which it travels. The formula for the law is $A = \epsilon c \cdot l$, where A is the absorbance (which has no units because it is a ratio), ϵ is the molar absorptivity or extinction coefficient ($L \text{ mol}^{-1} \text{ cm}^{-1}$), c is the absorbing species concentration (mol L^{-1}), and l is the sample's route length (cm).

According to this law, the sample's journey length and the absorbing species' concentration are directly correlated with the absorbance.

Benefits :

- High Sensitivity: Able to identify low amounts of chemicals;
- Non-Destructive: Does not modify or destroy the sample.

- Quantitative Analysis: Offers accurate numerical data regarding the specimen.

Drawbacks :

Interference: Other substances in the sample can interfere with the absorption measurements, leading to inaccurate results.

Sample Preparation: Some samples require extensive preparation, which can be time-consuming and may introduce potential for contamination.

Limited to Absorbing Species: This technique only works for substances that absorb light in the relevant wavelength range, limiting its applicability.

Quantitative Limitations: While it provides qualitative information effectively, quantitative analysis can be challenging due to variations in absorption efficiency.

Sensitivity to Environmental Conditions: Factors such as temperature and pressure can affect the accuracy of the measurements.

Interference: Inaccurate results may arise from other chemicals in the sample interfering with the absorption measurements.

Sample Preparation: Preparing some samples in-depth can take a lot of time and increase the risk of contamination.

Restricted to Absorbing Species: The usefulness of this technique is limited to compounds that absorb light within the relevant wavelength range.

Quantitative Limitations: Because absorption efficiency varies, quantitative analysis might be difficult even when it effectively offers qualitative information.

Sensitivity to Environmental Conditions: The precision of the measurements might be impacted by variables like pressure and temperature.

Limitations:

Interference: Measurements of absorption may be impacted by other compounds present in the sample.

Sample Preparation: A lot of preparation is necessary for some samples.

II. CONCLUSION :

Spectroscopy is a vital and adaptable analytical method that is important to many different branches of science. Through the examination of the relationship between matter and electromagnetic radiation, spectroscopy offers significant understanding of the makeup, arrangement, and characteristics of various substances.

Spectroscopy is a well-established, ancient technique with many layers and interfaces that require practice to become proficient in. But now that new photovoltaic technologies are emerging, the work is worthwhile. Impedance spectroscopy offers up new avenues for integration with other research fields, like materials science, for use in a new paradigm. Traditionally, it has been used in macroscopic settings to study the conductivity and electronic structure of samples internally, or the charge separation and optoelectronic properties of photovoltaic materials.

Future trends:

Miniaturization: Creation of field-use portable spectrophotometers.

Automation: Combining machine learning and artificial intelligence to automate data analysis.

Enhanced Sensitivity: Technological developments that enable the detection of even lower drug concentrations.

REFERENCE :

- [1]. Khan, ZuhaibAshfaq, Hafiz HusnainRazaSherazi, Mubashir Ali, Muhammad Ali Imran, Ikram Ur Rehman, and PrasunChakrabarti. "Designing a wind

- energy harvester for connected vehicles in green cities." *Energies* 14, no. 17 (2021): 5408.
- [2]. Gurdeep R. Chatwal, Anand Sharma *Instrumental method of Analysis* 5th edition, Himalaya publishing house PVT.LTD, 2002 pp 2.566-2.587.
- [3]. Amira H. Kamal, Samah F, et al. "A Review on UV Spectrophotometric Methods For Simultaneous Multicomponent Analysis ", *European Journal of Pharmaceutical And Medical Research*, 2016,3(2), 348-360.
- [4]. M. Patil, "Advantages and Disadvantages of UV Visible Spectroscopy", NOV 2018, <https://chrominfo.blogspot.com/2018/11/a-dvantages-and-disadvantages-of-uv.html?m=1>
- [5]. verma govinda, Dr.mishra Manish, "Development and optimization of UV -vis-spectroscopy -A Review, *world journal of pharmaceutical research*, 2018,7(11),1170-1180
- [6]. Skoog, D.A holler, F.J, nieman, D.A, *introduction to uv spectroscopy in principle of instrumental analysis*, 5th edition, cole publication, 2004
- [7]. *Basic uv spectroscopy theory, concepts and applications*, 7-10
- [8]. pateldiya, panchaldiya, patelkunj, "uv visible spectroscopy -A Review, *international journal of creative Research thought*, 2022,10(10),399-411
- [9]. Ganesh s. et al. "A review on advances in uv spectroscopy ". *Research j.science and tech* .2020,12(1),47-51
- [10]. priya, Archana G. "UV spectrophotometric method development and validation for amlodipine besylate in bulk and dosage form." *Asian journal of pharmaceutical analysis*, 2022,12(2),94-98
- [11]. Aneesimran mohammed, *Determinations of uv spectrophotometric method*, *International journal of current microbiology and applied science*, 2015,4(2),756-763
- [12]. savale Sagar, mahajan hitendra, *Asian journal of biomaterial research*, 2017,3(2):40-43
- [13]. Bikkasani, novel UV spectrophotometric method for simultaneous estimation of itraconazole, *world journal of pharmaceutical research*, 2021,10(4),1391-1454
- [14]. Rajyaalaxmi, *world journal of Pharmacy and Pharmaceutical science -a review on uv spectroscopy*, 2022,11(5):984-999
- [15]. morejanvi, *derivative UV spectroscopy*, *world journal of Pharmacy and pharmaceutical sciences*, 10(10),772-790
- [16]. sheteaarti.J, *recent advances in uv spectroscopy*, *International journal for Research trends and innovation*, 12(8):301-304
- [17]. Dhange Bhagyashrivishivasharan, *international journal of novel research and development, UV -visible Spectroscopy*, 2023,8(5):81-104
- [18]. Upadhayniraj, kumar Vijay, *International Research journal of environmental science, spectroscopic method for the detection of organophosphate pesticide current world environ*, 2013.8(2)
- [19]. Karthika B.R, mr.nishand V.M, DR.prasobh G.R, *infrared spectroscopy, international journal of research publications and review*, 2022,(3);526-552
- [20]. Bhagwat S.A, Gite A.V, khemnar S.y, phad S.H, Rokade V.G, *IR spectroscopy, international research journal of modernization in engineering technology and science*, 2024 6(3);4906-4913
- [21]. A Review article on :an overview on infrared spectroscopy, By -mr.karthik B.R
- [22]. kaiyyum A.bhalder, vipul M.patil. *study of nuclear magnetic resonance spectroscopy with application:-A comprehensive Review, international journal of creative Research thought*;824-835
- [23]. grant D.M, Harris R.K, Chichester, new york:-john wiley;1956. *encyclopedia of nuclear magnetic resonance*.
- [24]. prof.dharmadhikari.pooja .A, waghmare Santosh, *Mass spectroscopy -A Review, international journal of pharmaceutical research and applications*, 2022,7(3),410-416.
- [25]. MS yerlekar.ashwini. *Mass spectrometry - --A Review, international journal of pharmaceutical research*, 2014;42-45.